



Approved Signatories:

- | | | | |
|--------------------------|----------|-------------------------------------|------------|
| <input type="checkbox"/> | S. Eaton | <input type="checkbox"/> | B Stacey |
| <input type="checkbox"/> | D Hector | <input type="checkbox"/> | S Stratton |
| <input type="checkbox"/> | N Rand | <input checked="" type="checkbox"/> | S Telfer |
| <input type="checkbox"/> | B Davies | <input type="checkbox"/> | S Gray |

Signed:

Date of issue: 02 February 2021

Certificate Number: 5295

Customer Name and Address:

Scottish Government
 Water, Air, Soils and Flooding Division
 Environmental Quality Directorate
 Scottish Government
 Victoria Quay
 Edinburgh
 EH6 6QQ

Description:

Calibration factors for the air monitoring station(s) at
 Aberdeen City Council

Ricardo Energy & Environment ID:

ED11194/5295

Aberdeen City Council
 NOx analysers

Station	Date of Audit	Species	Analyser Serial no	Zero Response ¹	Zero uncertainty ppb	Calibration Factor ²	Factor uncertainty %	Converter eff. (%) ³
Aberdeen Anderson Drive	13 January 2021	NOx	697	1.7	2.6	1.1430	3.50	90.9
		NO		0.7		1.1598		
Aberdeen King St	13 January 2021	NOx	6785	3.3	2.6	1.1204	3.50	98.4
		NO		0.3		1.1141		
Aberdeen Market Street 2	12 January 2021	NOx	3507	3.2	2.5	1.0400	3.52	98.4
		NO		1.7		0.9886		

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Ricardo Energy & Environment
 18 Blythswood Square (2nd Floor),
 Glasgow,
 G2 4BG
 Tel: 01235 753205

Registered office
 Shoreham Technical Centre
 Shoreham-by-Sea
 West Sussex
 BN43 5FG
Registered in England No.
 08229264
VAT Registration No.
 GB 212 8365 24



CERTIFICATE OF CALIBRATION



Date of issue: 02 February 2021

Certificate Number: 5295

Ricardo Energy & Environment ID: ED11194/5295

PM10 analysers

Station	Date of audit	Analyser Serial no	Calculated kg^5	Uncertainty %	Total flow ⁴	Uncertainty %	Main flow	Uncertainty %
Aberdeen Anderson Drive	13 January 2021	1200c175870309	13167	1.0	15.97	2.2	2.89	2.2
Aberdeen Union Street Roadside	12 January 2021	1405A227711402	16880	1.0	13.12	2.2	4.44	2.2

PM2.5 analysers

Station	Date of audit	Analyser Serial no	Calculated kg^5	Uncertainty %	Total flow ⁴	Uncertainty %	Main flow	Uncertainty %
Aberdeen Union Street Roadside	12 January 2021	1405A227711402	15672	1.0	13.12	2.2	4.44	2.2

FIDAS analyser

Station	Date of audit	Analyser Serial no	Calculated kg^5	Uncertainty %	Total flow ⁴	Uncertainty %	Main flow	Uncertainty %
Aberdeen King St	13 January 2021	8374			4.20	2.2		2.2
Aberdeen Market Street 2	12 January 2021	6653			4.33	2.2		2.2





CERTIFICATE OF CALIBRATION



Page 3 of 3

Date of issue: 02 February 2021
Certificate Number: 5295
Ricardo Energy & Environment ID: ED11194/5295

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NO_x analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and k_0 (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are given in ppb (parts per billion) mole fractions or ppm (parts per million) mole fractions.

¹ The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

² The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported concentration units (ppb for NO, NO_x, SO₂, O₃ and ppm for CO. Where 1ppm = 1000ppb). It should be used in conjunction with the zero response. A corrected concentration is calculated using the following equation:

Concentration = F(Output - Zero Response)

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser

Zero Response = Zero Response provided on this certificate

³ Converter eff. is the measured efficiency of the NO₂ to NO converter within the oxides of nitrogen analyser under test.

⁴ The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min⁻¹, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

⁵ The calculated k_0 value (specifically for TEOM analysers) is the calculated k_0 spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified value of k_0 .

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.