



Approved Signatories:

- |                          |          |                                     |            |
|--------------------------|----------|-------------------------------------|------------|
| <input type="checkbox"/> | S. Eaton | <input type="checkbox"/>            | B Stacey   |
| <input type="checkbox"/> | D Hector | <input type="checkbox"/>            | S Stratton |
| <input type="checkbox"/> | N Rand   | <input checked="" type="checkbox"/> | S Telfer   |
| <input type="checkbox"/> | B Davies | <input type="checkbox"/>            | S Gray     |

Signed:

Date of issue: 24 February 2021

Certificate Number: 5276

Customer Name and Address:

Scottish Government  
Water, Air, Soils and Flooding Division  
Environmental Quality Directorate  
Scottish Government  
Victoria Quay  
Edinburgh  
EH6 6QQ

Description:

Calibration factors for the air monitoring station(s) at  
North Lanarkshire Council

Ricardo Energy & Environment ID:

ED11194/5276

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory

**Ricardo Energy & Environment**

18 Blythswood Square (2<sup>nd</sup> Floor),  
Glasgow,  
G2 4BG

Tel: 01235 753205

**Registered office**

Shoreham Technical Centre  
Shoreham-by-Sea  
West Sussex  
BN43 5FG

**Registered in England No.**

08229264

**VAT Registration No.**

GB 212 8365 24



CERTIFICATE OF CALIBRATION



Date of issue: 24 February 2021  
 Certificate Number: 5276  
 Ricardo Energy & Environment ID: ED11194/5276

North Lanarkshire Council  
 NOx analysers

Station	Date of Audit	Species	Analyser Serial no	Zero Response <sup>1</sup>	Zero uncertainty % <sup>2</sup>	Calibration Factor <sup>2</sup>	Factor uncertainty %	Converter eff. (%) <sup>3</sup>
N Lanarkshire Airdrie Kenilworth Drive	07 January 2021	NOx	9L2FE9GJ	-0.9	2.5	0.9686	3.50	98.9
		NO		-0.8	2.5	0.9758	3.50	
N Lanarkshire Chapelhall	06 January 2021	NOx	7NLHD0L8	-3.1	3.0	0.9882	3.50	102.0
		NO		-2.5	2.9	0.9918	3.50	
N Lanarkshire Coatbridge Sunnyside Road	06 January 2021	NOx	HUK14070019	0.1	2.7	1.2864	3.50	99.5
		NO		-0.6	2.7	1.3079	3.50	
N Lanarkshire Coatbridge Whifflet	12 February 2021	NOx	XDG8LYSO	0.3	2.7	1.1392	3.50	86.0
		NO		0.3	2.7	1.1594	3.50	
N Lanarkshire Croy	08 January 2021	NOx	AYKTCJU8	0.9	3.3	2.2538	3.50	101.8
		NO		-0.4	3.3	2.3256	3.50	
N Lanarkshire Motherwell	06 January 2021	NOx	YPB4FS4U	0.7	2.5	1.0631	3.50	98.3
		NO		0.0	2.6	1.0705	3.50	
N Lanarkshire Motherwell Adele Street	12 February 2021	NOx	EUGBA000	0.5	3.5	0.9938	3.50	100.7
		NO		0.8	3.2	0.9908	3.50	
N Lanarkshire Shawhead Coatbridge	05 January 2021	NOx	7NHSKHBC	-1.6	2.6	0.9783	3.50	99.6
		NO		-0.5	2.5	0.9859	3.50	
N Lanarkshire Kirkshaws	05 January 2021	NOx	HUK15020066	-0.8	2.6	1.0773	3.50	100.4
		NO		-0.6	2.6	1.0903	3.50	
N Lanarkshire Uddingston New Edinburgh Road	07 January 2021	NOx	-	-1.7	2.5	0.9536	3.50	103.2
		NO	(Inaccessible panel)	-0.9	2.5	0.9636	3.50	

PM10 analysers

Station	Date of audit	Analyser Serial no	Calculated ko <sup>3</sup>	Uncertainty %	Total flow <sup>4</sup>	Uncertainty %	Main flow	Uncertainty %
N Lanarkshire Airdrie Kenilworth Drive	07 January 2021	R11772			12.61	2.2		2.2
N Lanarkshire Coatbridge Sunnyside Road	06 January 2021	R11774			15.16	2.2		2.2
N Lanarkshire Uddingston New Edinburgh Road	07 January 2021	P18029			0.00	2.2		2.2

Fidas analysers

Station	Date of audit	Analyser Serial no	Calculated ko <sup>3</sup>	Uncertainty %	Total flow <sup>4</sup>	Uncertainty %	Main flow	Uncertainty %
N Lanarkshire Chapelhall	06 January 2021	8323			4.42	2.2		2.2
N Lanarkshire Coatbridge Whifflet	05 January 2021	12143			4.49	2.2		2.2
N Lanarkshire Croy	08 January 2021	9552			4.40	2.2		2.2
N Lanarkshire Kirkshaws	05 January 2021	9554			4.46	2.2		2.2
N Lanarkshire Motherwell	06 January 2021	9551			4.44	2.2		2.2
N Lanarkshire Motherwell Adele Street	07 January 2021	9553			4.41	2.2		2.2
N Lanarkshire Shawhead Coatbridge	05 January 2021	9550			4.45	2.2		2.2



## CERTIFICATE OF CALIBRATION



Page 3 of 3

Date of issue: 24 February 2021  
Certificate Number: 5276  
Ricardo Energy & Environment ID: ED11194/5276

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NO<sub>x</sub> analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and *k<sub>o</sub>* (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are given in ppb (parts per billion) mole fractions or ppm (parts per million) mole fractions.

<sup>1</sup> The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

<sup>2</sup> The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported concentration units (ppb for NO, NO<sub>x</sub>, SO<sub>2</sub>, O<sub>3</sub> and ppm for CO. Where 1ppm = 1000ppb). It should be used in conjunction with the zero response. A corrected concentration is calculated using the following equation:

$$\text{Concentration} = F(\text{Output} - \text{Zero Response})$$

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser

Zero Response = Zero Response provided on this certificate

<sup>3</sup> Converter eff. is the measured efficiency of the NO<sub>2</sub> to NO converter within the oxides of nitrogen analyser under test.

<sup>4</sup> The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min<sup>-1</sup>, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

<sup>5</sup> The calculated *k<sub>o</sub>* value (specifically for TEOM analysers) is the calculated *k<sub>o</sub>* spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified value of *k<sub>o</sub>*.

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.