





0401

CERTIFICATE OF CALIBRATION

Ricardo Energy & Environment, 18 Blythswood Square, Glasgow, G2 4BG Telephone 01235 753642

Authorised Signatories:

D Hector ✓

S Stratton

Signed:

Date of Issue: 18th July 2017

Certificate Number: 3745

Page 1 of 1

Customer Name and Address: Scottish Government

> Water, Air, Soils and Flooding Division **Environmental Quality Directorate**

Scottish Government

Victoria Quay Edinburgh EH6 6QQ

Description:

Calibration factors for Angus Council's Forfar air monitoring station.

Site / Date Test Carried Out	Species	Analyser Serial No.	Parameter	Specified Value	Measured Value	Deviation %	Uncertainty %
			Main Flow ¹	3.00	2.97	-0.8	2.25
Forfar Glamis Road	FDMS	26569	Aux Flow ¹	13.67			
10 th August 2016	PM ₁₀	ļ	Total Flow	16.67	15.48	-7.1	2.25
			k_0^2	14313	14225	-0.6	1.00

¹The measured main flow rate (where applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured aux flow rate (where applicable) is the flow rate through the bypass tubing of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are I.min-1. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

The calibration results shaded are those that fall out with our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95% The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. Ricardo Energy & Environment is a trading name of Ricardo-AEA Ltd.

Ricardo Energy & Environment

Registered office Shoreham Technical Centre Shoreham-by-Sea West Sussex BN43 5FG

Gemini Building, Fermi Avenue, Harwell, **OX11 0QR**

Head Office

Registered in England No.

08229264

+44 (0)1235 753 000

VAT Registration No. GB 212 8365 24

Delivering Excellence Through Innovation & Technology

²The calculated k0 value (TEOM analysers only) is the calculated k0 spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified k0 value.