

CERTIFICATE OF CALIBRATION

Ricardo Energy and Environment, Gemini Building, Fermi Avenue Harwell, Didcot, Oxfordshire OX11 OQR. Telephone 01235 753692



Page 1 of 3

Approved Signatories:		S. Eaton D Hector N Rand B Davies	☐ B Stacey ☐ S Stratton ☐ S Telfer ☑ S Gray			
Signed: Date of issue:	29 Apr 19					
Certificate Number:	29 Apr 19 4486					
Customer Name and Address:		Scottish Government Water, Air, Soils and Flooding Dir Environmental Quality Directora Scottish Government Victoria Quay Edinburgh EH6 6QQ				
Description:		Calibration factors for the air monitoring station(s) at South Lanarkshire Council				
Ricardo Energy & Environment ID:		ED61598/4486				
The reported expanded uncertainties are based of level of confidence of approximately 95% The un requirements. This certificate is issued in accordance with the last Service. It provides traceability of measurement National Physical Laboratory or other recognised than in full, except with the prior written approv	certainty evaluation has le aboratory accreditation re to the SI system of units a I national metrology instit al of the issuing laborator	peen carried out in accordance with UKAS equirements of the United Kingdom Accred and/or to units of measurement realised at tutes. This certificate may not be reproduce y	itation the			
Ricardo Energy & Environment Head Office Gemini Building, Fermi Avenue, Harwell, Oxon OX11 0QR Tel: +44 (0)1235 753 000	Registered off Shoreham Tect Shoreham-by-S West Sussex BN43 5FG Registered in 1 08229264 VAT Registrat GB 212 8365 2-	anical Centre England No.				

ee.**ricardo**.com



CERTIFICATE OF CALIBRATION



Page 2 of 3

Date of issue: 29 Apr 19

Certificate Number: 4486

ED61598/4486 Ricardo Energy & Environment ID:

South Lanarkshire Council NOx analysers

Station	Date of Audit	Species	Analyser Serial no	Zero Response ¹	Zero uncertainty ppb	Calibration Factor ²	Factor uncertainty %	Converter eff. (%) ³
South Lanarkshire Blantyre	20-Mar-19	NOx	18-0740	2.0	2.6	1.1566	3.50	100.0
		NO		0.0	2.6	1.1318	3.50	
South Lanarkshire Cambuslang	04-Dec-18	NOx	1152590008	0.1	2.5	0.9445	4.09	98.2
		NO		-0.4	2.5	0.9354	4.05	
South Lanarkshire East Kilbride	04-Dec-18	NOx	7460075	0.1	2.6	1.2088	3.54	101.5
		NO		-0.3	2.6	1.1996	3.50	
South Lanarkshire Hamilton	04-Dec-18	NOx	7460073	-6.3	2.5	1.0216	3.50	101.5
		NO		-0.9	2.5	1.0320	3.50	
South Lanarkshire Raith Interchange 2	06-Dec-18	NOx	102200001	-0.4	2.5	1.0057	4.08	98.8
		NO		-0.4	2.5	0.9991	4.12	
South Lanarkshire Rutherglen	07-Dec-18	NOx	7460076	-0.9	2.5	0.9644	3.81	98.6
		NO		-0.8	2.5	0.9583	3.86	
South Lanarkshire Uddingston 0	06-Dec-18	NOx	10020068	-1.4	2.5	0.9716	3.64	100.0
		NO		-1.4	2.5	0.9673	4.28	
South Lanarkshire Lanark 06-I	06-Dec-18	NOx	10020067	-0.3	2.5	0.9780	3.50	98.4
		NO		-0.2	2.5	0.9760	3.50	

PM10 analysers

Station	Date of audit	Analyser Serial no	Calculated ko Uncertainty %	Total flow	Uncertainty %	Main flow	Uncertainty %
South Lanarkshire Blantyre	20-Mar-19	9750		4.61	2.2		2.2
South Lanarkshire Cambuslang	04-Dec-18	8256		4.75	2.2		2.2
South Lanarkshire East Kilbride	04-Dec-18	8257		4.86	2.2		2.2
South Lanarkshire Hamilton	04-Dec-18	8258		5.16	2.2		2.2
South Lanarkshire Raith Interchange 2	06-Dec-18	9377		4.51	2.2		2.2
South Lanarkshire Rutherglen	07-Dec-18	8140		4.71	2.2		2.2
South Lanarkshire Uddingston	06-Dec-18	6247		4.57	2.2		2.2
South Lanarkshire Lanark	06-Dec-18	6248		4.75	2.2		2.2

PM2.5 analysers

1 WIZ.3 dildiyacia							
Station	Date of audit	Analyser Serial no	Calculated ko Uncertainty %	Total flow	Uncertainty %	Main flow	Uncertainty %
South Lanarkshire Blantyre	20-Mar-19	9750		4.61	2.2		2.2
South Lanarkshire Cambuslang	21-Mar-19	8256		4.75	2.2		2.2
South Lanarkshire East Kilbride	22-Mar-19	8257		4.86	2.2		2.2
South Lanarkshire Hamilton	23-Mar-19	8258		5.16	2.2		2.2
South Lanarkshire Raith Interchange 2	24-Mar-19	9377		4.51	2.2		2.2
South Lanarkshire Rutherglen	25-Mar-19	8140		4.71	2.2		2.2
South Lanarkshire Uddingston	26-Mar-19	6247		4.57	2.2		2.2
South Lanarkshire Lanark	27-Mar-19	6248		4.75	2.2		2.2

ee.**ricardo**.com



CERTIFICATE OF CALIBRATION



Page 3 of 3

Date of issue:

29 Apr 19

- ¹ The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.
- ² The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported concentration units (ppb for NO, NOx, SO₂, O₃ and ppm for CO. Where 1ppm = 1000ppb). It should be used in conjunction with the zero response. A corrected concentration is calculated using the following equation:

Concentration = F(Output - Zero Response)

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser Zero Response = Zero Response provided on this certificate

- ³ Converter eff. is the measured efficiency of the NO₂ to NO converter within the oxides of nitrogen analyser under test.
- ⁴ The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min⁻¹, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.
- ⁵ The calculated ko value (specifically for TEOM analysers) is the calculated ko spring constant based on tests undertaken with filters of known weight.

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.

ee.ricardo.com