



## **CERTIFICATE OF CALIBRATION**

Ricardo Energy & Environment 18 Blythswood Square, Glasgow, G2 4BG





Page 1 of 3

Approved Signatories:	□ S. Eaton □ B Stacey □ S Copsey □ S Stratton □ N Rand □ S Telfer □ B Davies □ S Gray □ D Lane □ T Green □ A Nash
Signed:	3450
Date of issue:	01 April 2025
Certificate Number:	7389
Customer Name and Address:	Scottish Government Water, Air, Soils and Flooding Division Environmental Quality Directorate Scottish Government Victoria Quay Edinburgh EH6 6QQ
Description:	Calibration factors for the air monitoring station(s) at North Lanarkshire Council
Ricardo Energy & Environment ID:	Ed19050/7389
level of confidence of approximately 95% The requirements.  This certificate is issued in accordance with th Service. It provides traceability of measurements National Physical Laboratory or other recognithan in full, except with the prior written approximately approx	Registered office Shoreham Technical Centre Shoreham-by-Sea West Sussex BN43 5FG Registered in England No.
Tel: 01235 753205	Registered in England No. 08229264

VAT Registration No. GB 212 8365 24

ee.**ricardo**.com



# **CERTIFICATE OF CALIBRATION**



Page 2 of 3

Date of issue: 01 April 2025

Certificate Number: 7389

Ricardo Energy & Environment ID: Ed19050/7389

#### North Lanarkshire Council

NOx analysers

NOX dilalysers									
Station	Date of Audit	Species	Analyser Serial no	Zero Response <sup>1</sup>	Zero uncertainty nmol/mol	Calibration Factor <sup>2</sup>	Factor uncertainty %	Converter eff. (%) <sup>3</sup>	Converter uncertainty (%)
N Lanarkshire Chapelhall	07 January 2025	NOx	23-0185	3.1	2.7	1.3639	3.59	100.2 (272 nmol/mol)	3.9
		NO		2.3	2.7	1.3623	3.50	99.7 (130 nmol/mol)	3.9
N Lanarkshire Gartcosh	06 January 2025	NOx	23-0158	0.1	2.5	0.8689	3.50	98.7 (233 nmol/mol)	0.2
		NO		0.2	2.5	0.8687	3.50	99.5 (106 nmol/mol)	0.2
N Lanarkshire Kirkshaws	20 March 2025	NOx	23-0155	0.6	2.5	1.0481	3.50	99.9 (266 nmol/mol)	0.5
		NO		0.9	2.5	1.0425	3.50	99.5 (127 nmol/mol)	0.5
N Lanarkshire Ravenscraig Plantatior	09 January 2025	NOx	23-0552	0.6	2.5	1.0451	3.50	99.3 (272 nmol/mol)	1.3
		NO		1.0	2.5	1.0457	3.50	98.9 (133 nmol/mol)	1.3
N Lanarkshire Shawhead Coatbridge	07 January 2025	NOx	23-0160	-0.1	2.5	0.9979	3.54	99.1 (228 nmol/mol)	2.7
		NO		0.2	2.5	0.9931	3.61	98.8 (105 nmol/mol)	2.7

#### Fidas analysers

11445 4114175615							
Station	Date of audit	Analyser Serial no	Zero (µg/m³)	Caldust channel deviation	Total flow⁴	Uncertainty %	Deviation %
N Lanarkshire Kirkshaws 2	07 January 2025	9554	0	1.30	5.35	2.25	11.55
N Lanarkshire Motherwell	06 January 2025	9551	0	0.98	5.26	2.25	9.58
N Lanarkshire Ravenscraig Plantation	09 January 2025	12143	0	1.22	4.95	2.25	3.16
N Lanarkshire Shawhead Coatbridge	07 January 2025	9550	0	0.42	5.12	2.25	6.60

ee.**ricardo**.com



### **CERTIFICATE OF CALIBRATION**



Page 3 of 3

Date of issue: 01 April 2025

Certificate Number: 7389

Ricardo Energy & Environment ID: Ed19050/7389

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NOx analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are reported in composition units of nmol/mol or  $\mu$ mol/mol.

composition = F(Output - Zero Response)

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser

Zero Response = Zero Response provided on this certificate

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.

\*\*\*\*\*END OF CERTIFICATE\*\*\*\*\*

ee.**ricardo**.com

<sup>&</sup>lt;sup>1</sup> The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

 $<sup>^2</sup>$  The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported composition units (nmol/mol for NO, NOx, SOz, O₃ and µmol/mol for CO). It should be used in conjunction with the zero response. A corrected composition is calculated using the following equation:

 $<sup>^{3}</sup>$  Converter eff. is the measured efficiency of the NO2 to NO converter within the oxides of nitrogen analyser under test.

<sup>&</sup>lt;sup>4</sup> The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured aux flow rate (where this is applicable) is the flow rate through the bypass tubing of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min-1, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

<sup>&</sup>lt;sup>5</sup> The calculated k0 value (specifically for TEOM analysers) is the calculated k0 spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified value of k0