



CERTIFICATE OF CALIBRATION

Ricardo Energy & Environment 18 Blythswood Square, Glasgow, G2 4BG





Page 1 of 3

Approved Signatories:	□ S. Eaton □ B Stacey □ S Copsey □ S Stratton □ N Rand □ S Telfer □ B Davies □ S Gray □ D Lane □ F. Elmer □ A Nash	
Signed:	346	
Date of issue:	04 September 2025	
Certificate Number:	7610	
Customer Name and Address:	Scottish Government Water, Air, Soils and Flooding Division Environmental Quality Directorate Scottish Government Victoria Quay Edinburgh EH6 6QQ	
Description:	Calibration factors for the air monitoring station(s) a North Lanarkshire Council	ıt
Ricardo Energy & Environment ID:	ED19050 /7610	
level of confidence of approximately 95% The requirements. This certificate is issued in accordance with th Service. It provides traceability of measurements.	d on a standard uncertainty multiplied by a coverage factor k=2 providing a uncertainty evaluation has been carried out in accordance with UKAS e laboratory accreditation requirements of the United Kingdom Accreditation nt to the SI system of units and/or to units of measurement realised at the sed national metrology institutes. This certificate may not be reproduced other oval of the issuing laboratory Registered office	
Ricardo Energy & Environment 18 Blythswood Square (2 nd Floor), Glasgow, G2 4BG Tel: 01235 753205	Shoreham Technical Centre Shoreham-by-Sea West Sussex BN43 5FG Registered in England No. 08229264	

VAT Registration No. GB 212 8365 24

ee.**ricardo**.com



CERTIFICATE OF CALIBRATION



Page 2 of 3

Date of issue: 04 September 2025

Certificate Number: 7610

Ricardo Energy & Environment ID: ED19050 /7610

North Lanarkshire Council

NOx analysers

INOX dilalysers									
Station	Date of Audit	Species	Analyser Serial no	Zero Response ¹	Zero uncertainty nmol/mol	Calibration Factor ²	Factor uncertainty %	Converter eff. (%) ³	Converter uncertainty (%)
N Lanarkshire Chapelhall	23 June 2025	NOx	23-0185	-10.7	2.8	1.1257	3.50	99.6 (267 nmol/mol)	1.7
		NO		0.1	2.6	1.1240	3.50	99 (127 nmol/mol)	1.7
N Lanarkshire Gartcosh	20 June 2025	NOx	23-0158	3.6	2.5	1.0780	3.50	100.4 (275 nmol/mol)	0.5
		NO		2.1	2.6	1.0768	3.50	101.6 (134 nmol/mol)	0.5
N Lanarkshire Kirkshaws	25 June 2025	NOx	23-0155	1.1	2.5	1.0561	3.50	100.1 (297 nmol/mol)	0.3
		NO		1.1	2.5	1.0513	3.50	99.8 (152 nmol/mol)	0.3
N Lanarkshire Ravenscraig	26 June 2025	NOx	23-0552	6.4	2.5	1.0205	3.50	100.4 (288 nmol/mol)	1.3
Plantation Road		NO		6.9	2.5	1.0190	3.50	98.6 (171 nmol/mol)	1.3
N Lanarkshire Shawhead Coatbridge	25 June 2025	NOx	23-0195	0.5	2.6	1.1238	3.50	100.6 (287 nmol/mol)	1.5
		NO		0.4	2.6	1.1203	3.50	100.2 (148 nmol/mol)	1.5

Fidas analysers

Station	Date of audit	Analyser Serial no	Zero (µg/m³)	Caldust channel deviation	Total flow⁴	Uncertaint y %	Deviation %
N Lanarkshire Kirkshaws	25 June 2025	9554	0	0.70	4.82	2.25	0.51
N Lanarkshire Motherwell	23 June 2025	9551	0	1.99	4.86	2.25	1.20
N Lanarkshire Ravenscraig Plantation Road	26 June 2025	12143	0	1.28	4.74	2.25	-1.20
N Lanarkshire Shawhead Coatbridge	25 June 2025	9550	0	1.66	4.78	2.25	-0.32

ee.**ricardo**.com



CERTIFICATE OF CALIBRATION



Page 3 of 3

Date of issue: 04 September 2025

Certificate Number: 7610

Ricardo Energy & Environment ID: ED19050 /7610

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NOx analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are reported in concentration units of nmol/mol or µmol/mol.

Concentration = F(Output - Zero Response)

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser

Zero Response = Zero Response provided on this certificate

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.

*****END OF CERTIFICATE*****

ee.**ricardo**.com

¹ The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

² The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported concentration units (nmol/mol for NO, NOx, SO2, O3 and µmol/mol for CO). It should be used in conjunction with the zero response. A corrected concentration is calculated using the following equation:

 $^{^{3}}$ Converter eff. is the measured efficiency of the NO2 to NO converter within the oxides of nitrogen analyser under test.

⁴ The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured aux flow rate (where this is applicable) is the flow rate through the bypass tubing of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min-1, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

⁵ The calculated k0 value (specifically for TEOM analysers) is the calculated k0 spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified value of k0.