



Approved Signatories:

- | | | | |
|--------------------------|----------|-------------------------------------|------------|
| <input type="checkbox"/> | S. Eaton | <input type="checkbox"/> | B Stacey |
| <input type="checkbox"/> | D Hector | <input type="checkbox"/> | S Stratton |
| <input type="checkbox"/> | N Rand | <input checked="" type="checkbox"/> | S Telfer |
| <input type="checkbox"/> | B Davies | <input type="checkbox"/> | S Gray |

Signed:

Date of issue:

23 March 2023

Certificate Number:

6214

Customer Name and Address:

North Lanarkshire Council
Municipal Buildings
Kildonan Street
Coatbridge
ML5 3LF

Description:

Calibration factors for the air monitoring station(s) at
North Lanarkshire Council

Ricardo Energy & Environment ID:

ED15141/6214

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k=2$ providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Ricardo Energy & Environment

18 Blythswood Square (2nd Floor),
Glasgow,
G2 4BG

Tel: 01235 753205

Registered office

Shoreham Technical Centre
Shoreham-by-Sea
West Sussex
BN43 5FG

Registered in England No.

08229264

VAT Registration No.

GB 212 8365 24



CERTIFICATE OF CALIBRATION



Date of issue: 23 March 2023
 Certificate Number: 6214
 Ricardo Energy & Environment ID: ED15141/6214

North Lanarkshire Council
 NOx analysers

Station	Date of Audit	Species	Analyser Serial no	Zero Response ¹	Zero uncertainty nmol/mol	Calibration Factor ²	Factor uncertainty %	Converter eff. (%) ³
N Lanarkshire Coatbridge Whifflet A725	16 January 2023	NOx	21-1547	-3.0	9.4	1.0470	3.95	100.4 (247nmol/mol)
		NO		0.0	7.2	1.0493	3.63	101.9 (110nmol/mol)
N Lanarkshire Airdrie Kenilworth Drive	18 January 2023	NOx	HUK14100092	-1.7	2.5	0.9905	3.50	99.5 (252nmol/mol)
		NO		-0.7	2.5	1.0030	3.50	100.6 (109nmol/mol)
N Lanarkshire Motherwell Adele Street	17 January 2023	NOx	21-0951	-2.0	2.9	1.0241	3.50	100.4 (270nmol/mol)
		NO		0.0	2.7	1.0377	3.61	102 (153nmol/mol)
N Lanarkshire Motherwell	17 January 2023	NOx	YPB4FS4U	0.0	2.7	1.2854	3.50	99.1 (269nmol/mol)
		NO		0.1	2.7	1.3131	3.50	99.1 (117nmol/mol)
N Lanarkshire Ravenscraig Plantation Road	19 January 2023	NOx	XDGBLYS0	-0.5	2.6	1.1630	3.50	99.2 (243nmol/mol)
		NO		-0.1	2.6	1.1721	3.50	99.8 (94nmol/mol)
N Lanarkshire Uddingston New Edinburgh Road	17 January 2023	NOx	Serial number obstructed	-2.0	2.7	0.9859	3.50	102.6 (241nmol/mol)
		NO		-0.8	2.8	0.9977	3.50	102.4 (107nmol/mol)

PM10 analysers

Station	Date of audit	Analyser Serial no	Calculated ko ²	Uncertainty %	Total flow ⁴	Uncertainty %	Main flow	Uncertainty %
N Lanarkshire Airdrie Kenilworth Drive	18 January 2023	R11772			15.91	2.2		2.2

Fidas analysers

Station	Date of audit	Analyser Serial no	Calculated ko ³	Uncertainty %	Total flow ⁴	Uncertainty %	Main flow	Uncertainty %
N Lanarkshire Coatbridge Whifflet A725	16 January 2023	14213			5.23	2.2		2.2
N Lanarkshire Motherwell Adele Street	17 January 2023	9553			5.21	2.2		2.2
N Lanarkshire Uddingston New Edinburgh Road	17 January 2023	158863			4.92	2.2		2.2





CERTIFICATE OF CALIBRATION



Date of issue: 23 March 2023
Certificate Number: 6214
Ricardo Energy & Environment ID: ED15141/6214

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NOx analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and k0 (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are reported in concentration units of nmol/mol or µmol/mol.

1 The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

2 The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported concentration units (nmol/mol for NO, NOx, SO2, O3 and µmol/mol for CO). It should be used in conjunction with the zero response. A corrected concentration is calculated using the following equation:

Concentration = F(Output - Zero Response)
Where F = Calibration Factor provided on this certificate
Output = Reading on the data logging system of the analyser
Zero Response = Zero Response provided on this certificate

3 Converter eff. is the measured efficiency of the NO2 to NO converter within the oxides of nitrogen analyser under test.

4 The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured aux flow rate (where this is applicable) is the flow rate through the bypass tubing of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min-1, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

5 The calculated k0 value (specifically for TEOM analysers) is the calculated k0 spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified value of k0.

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.

