



### **CERTIFICATE OF CALIBRATION**

Ricardo Energy & Environment 18 Blythswood Square, Glasgow, G2 4BG





Page 1 of 3

Approved Signatories:	□ S. Eaton □ B Stacey □ S Copsey □ S Stratton □ N Rand □ S Telfer □ B Davies □ S Gray □ D Lane □ T Green □ A Nash						
Signed:	Age 60						
Date of issue:	24 September 2024						
Certificate Number:	7008						
Customer Name and Address:	Scottish Government Water, Air, Soils and Flooding Division Environmental Quality Directorate Scottish Government Victoria Quay Edinburgh EH6 6QQ						
Description:	Calibration factors for the air monitoring station(s) at South Lanarkshire Council						
Ricardo Energy & Environment ID:	ED19050 / 7008						
The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95% The uncertainty evaluation has been carried out in accordance with UKAS requirements.  This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory							
Ricardo Energy & Environment  18 Blythswood Square (2nd Floor), Glasgow, G2 4BG  Tel: 01235 753205	Registered office Shoreham Technical Centre Shoreham-by-Sea West Sussex BN43 5FG  Registered in England No. 08229264						

VAT Registration No. GB 212 8365 24

ee.**ricardo**.com



# **CERTIFICATE OF CALIBRATION**



Page 2 of 3

Date of issue: 24 September 2024

Certificate Number: 7008

Ricardo Energy & Environment ID: ED19050 / 7008

## South Lanarkshire Council

NOx analysers

Station	Date of Audit	Species	Analyser Serial no	Zero Response <sup>1</sup>	Zero uncertainty nmol/mol	Calibration Factor <sup>2</sup>	Factor uncertainty %	Converter eff. (%) <sup>3</sup>	Converter uncertainty (%)
South Lanarkshire Blantyre	04 June 2024	NOx	18-0740	0.0	2.5	0.8621	3.50	99.3 (253nmol/mol)	8.0
		NO		0.0	2.6	0.8537	3.50	100 (119nmol/mol)	0.8
South Lanarkshire East Kilbride	06 June 2024	NOx	CM07460075	0.9	2.6	1.1085	3.50	100 (251nmol/mol)	1.3
		NO		0.7	2.6	1.1111	3.50	96.5 (123nmol/mol)	1.3
South Lanarkshire Lanark	14 August 2024	NOx	CM10020067	0.6	2.5	0.9844	3.50	100 (222nmol/mol)	1.8
		NO		0.4	2.5	0.9860	3.50	99.2 (124nmol/mol)	1.8
South Lanarkshire Raith Interchange	03 June 2024	NOx	CM07460076	-0.3	2.5	0.9744	3.57	98.8 (251nmol/mol)	1.3
2		NO		-0.2	2.5	0.9786	3.50	99.2 (118nmol/mol)	1.3
South Lanarkshire Rutherglen	03 June 2024	NOx	CM10220001	2.2	2.6	1.0862	3.50	97.1 (259nmol/mol)	0.6
		NO		2.6	2.6	1.0901	3.50	98.3 (125nmol/mol)	0.6

#### Fidas analysers

Trado arialysers							
Station	Date of audit	Analyser Serial no	Zero (µg/m³)	Caldust channel deviation	Total flow⁴	Uncertaint y %	Deviation %
South Lanarkshire Cambuslang	04 June 2024	8256	0	0.70	5.09	2.25	5.94
South Lanarkshire East Kilbride	16 August 2024	8257	0	1.60	5.03	2.25	4.71
South Lanarkshire Lanark 2	13 August 2024	6248	0	1.18	5.35	2.25	11.39
South Lanarkshire Raith Interchange	03 June 2024	9719	0	1.40	4.90	2.25	2.18
South Lanarkshire Rutherglen	03 June 2024	8140	0	0.43	4.84	2.25	0.77

ee.**ricardo**.com



#### **CERTIFICATE OF CALIBRATION**



Page 3 of 3

Date of issue: 24 September 2024

Certificate Number: 7008

Ricardo Energy & Environment ID: ED19050 / 7008

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NOx analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and k0 (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are reported in composition units of nmol/mol or  $\mu$ mol/mol.

composition = F(Output - Zero Response)

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser

Zero Response = Zero Response provided on this certificate

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.

\*\*\*\*\*END OF CERTIFICATE\*\*\*\*\*

ee.ricardo.com

<sup>&</sup>lt;sup>1</sup> The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

<sup>&</sup>lt;sup>2</sup> The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported composition units (nmol/mol for NO, NOx SO<sub>2</sub>, O<sub>3</sub> and µmol/mol for CO). It should be used in conjunction with the zero response. A corrected composition is calculated using the following equation:

<sup>&</sup>lt;sup>3</sup> Converter eff. is the measured efficiency of the NO2 to NO converter within the oxides of nitrogen analyser under test.

<sup>&</sup>lt;sup>4</sup> The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured aux flow rate (where this is applicable) is the flow rate through the bypass tubing of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min-1, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

<sup>&</sup>lt;sup>5</sup> The calculated k0 value (specifically for TEOM analysers) is the calculated k0 spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified value of k0.