



CERTIFICATE OF CALIBRATION

Ricardo Energy and Environment, Gemini Building, Fermi Avenue Harwell,
Didcot, Oxfordshire OX11 0QR. Telephone 01235 753692



Approved Signatories:

- | | |
|-----------------------------------|--|
| <input type="checkbox"/> S. Eaton | <input type="checkbox"/> B Stacey |
| <input type="checkbox"/> D Hector | <input type="checkbox"/> S Stratton |
| <input type="checkbox"/> N Rand | <input type="checkbox"/> S Telfer |
| <input type="checkbox"/> B Davies | <input checked="" type="checkbox"/> S Gray |

Signed:

Date of issue:

09 May 19

Certificate Number:

4507

Customer Name and Address:

Scottish Government
Water, Air, Soils and Flooding Division
Environmental Quality Directorate
Scottish Government
Victoria Quay
Edinburgh
EH6 6QQ

Description:

Calibration factors for the air monitoring station(s) at
Glasgow City Council

Ricardo Energy & Environment ID:

ED61598/4507

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95% The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory

Ricardo Energy & Environment

Head Office
Gemini Building,
Fermi Avenue,
Harwell,
Oxon
OX11 0QR

Tel: +44 (0)1235 753 000

Registered office

Shoreham Technical Centre
Shoreham-by-Sea
West Sussex
BN43 5FG

Registered in England No.
08229264

VAT Registration No.
GB 212 8365 24



CERTIFICATE OF CALIBRATION



Date of issue: 09 May 19
 Certificate Number: 4507
 Ricardo Energy & Environment ID: ED61598/4507

Glasgow City Council
 NOx analysers

Station	Date of Audit	Species	Analyser Serial no	Zero Response ¹	Zero uncertainty ppb	Calibration Factor ²	Factor uncertainty %	Converter eff. (%) ³
Glasgow Anderston	21-Jun-18	NOx	529b-229	1.0	2.5	1.0247	3.50	97.9
		NO		0.0	2.5	1.0294	3.50	
Glasgow Burgher St	20-Jun-18	NOx	cm11050006	-0.1	2.5	0.9325	3.50	98.8
		NO		-0.4	2.5	0.9278	3.50	
Glasgow Byres Road	20-Jun-18	NOx	4156	-13.3	2.5	1.0002	3.50	99.1
		NO		-7.5	2.5	1.0057	3.50	
Glasgow Dumbarton Road	20-Jun-18	NOx	4154	-6.7	2.4	0.7153	3.50	100.6
		NO		-6.7	2.4	0.6917	3.50	
Glasgow Waulkmillglen Reservoir	22-Jun-18	NOx	4155	-4.6	2.5	0.9841	3.50	100.4
		NO		-1.2	2.5	0.9861	3.50	

NO
 NOx
 NO

PM10 analysers

Station	Date of audit	Analyser Serial no	Calculated ko	Uncertainty %	Total flow	Uncertainty %	Main flow	Uncertainty %
Glasgow Abercromby Street	20-Jun-18	1200c191870611	15912	1.0	16.13	2.2	2.97	2.2
Glasgow Anderston	21-Jun-18	1200c187880603	13826	1.0	16.48	2.2	3.03	2.2
Glasgow Burgher St	20-Jun-18	1200c149419903	16364	1.0	15.80	2.2	2.85	2.2
Glasgow Byres Road	20-Jun-18	8734			4.15	2.2		2.2
Glasgow Dumbarton Road	27-Aug-18	8735			4.69	2.3		2.2
Glasgow Nithsdale Road	22-Jun-18	6249			4.24	2.2		2.2
Glasgow Waulkmillglen Reservoir	22-Jun-18	8736			4.39	2.2		2.2

PM2.5 analysers

Station	Date of audit	Analyser Serial no	Calculated ko	Uncertainty %	Total flow	Uncertainty %	Main flow	Uncertainty %
Glasgow Byres Road	20-Jun-18	8734			4.15	2.2		2.2
Glasgow Dumbarton Road	27-Aug-18	8735			4.69	2.3		2.2
Glasgow Nithsdale Road	22-Jun-18	6249			4.24	2.2		2.2
Glasgow Waulkmillglen Reservoir	22-Jun-18	8736			4.39	2.2		2.2





CERTIFICATE OF CALIBRATION



Date of issue: 09 May 19
Certificate Number: 4507
Ricardo Energy & Environment ID: ED61598/4507

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NO_x analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and k_0 (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are given in ppb (parts per billion) mole fractions or ppm (parts per million) mole fractions.

¹ The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

² The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported concentration units (ppb for NO, NO_x, SO₂, O₃ and ppm for CO. Where 1ppm = 1000ppb). It should be used in conjunction with the zero response. A corrected concentration is calculated using the following equation:

$$\text{Concentration} = F(\text{Output} - \text{Zero Response})$$

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser

Zero Response = Zero Response provided on this certificate

³ Converter eff. is the measured efficiency of the NO₂ to NO converter within the oxides of nitrogen analyser under test.

⁴ The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min⁻¹, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

⁵ The calculated k_0 value (specifically for TEOM analysers) is the calculated k_0 spring constant based on tests undertaken with filters of known weight.

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.