



0401

# CERTIFICATE OF CALIBRATION

Ricardo Energy and Environment, Gemini Building, Fermi Avenue Harwell,  
Didcot, Oxfordshire OX11 0QR. Telephone 01235 753692



Approved Signatories:

- S. Eaton
- D Hector
- N Rand
- B Davies

- B Stacey
- S Stratton
- S Telfer
- S Gray

Signed:

Date of issue:

20 May 20

Certificate Number:

4956

Customer Name and Address:

Scottish Government  
Water, Air, Soils and Flooding Division  
Environmental Quality Directorate  
Scottish Government  
Victoria Quay  
Edinburgh  
EH6 6QQ

Description:

Calibration factors for the air monitoring station(s) at  
North Lanarkshire Council

Ricardo Energy & Environment ID:

ED11194 / 4956

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory

**Ricardo Energy & Environment**  
18 Blythswood Square (2<sup>nd</sup> Floor),  
Glasgow,  
G2 4BG  
Tel: 01235 753205

**Registered office**  
Shoreham Technical Centre  
Shoreham-by-Sea  
West Sussex  
BN43 5FG  
**Registered in England No.**  
08229264  
**VAT Registration No.**  
GB 212 8365 24





# CERTIFICATE OF CALIBRATION



Date of issue: 20 May 20  
 Certificate Number: 4956  
 Ricardo Energy & Environment ID: ED11194 / 4956

North Lanarkshire Council  
 NOx analysers

Station	Date of Audit	Species	Analyser Serial no	Zero Response <sup>1</sup>	Zero uncertainty ppb	Calibration Factor <sup>2</sup>	Factor uncertainty %	Converter eff. (%) <sup>3</sup>
N Lanarkshire Chapelhall	18-Jul	NOx	7nlhd0l8	-0.3	2.5	0.9786	5.17	99.7
		NO		-0.2	2.5	1.0132	4.59	
N Lanarkshire Coatbridge Whifflet	19-Jul	NOx	XDG8LYS0	-0.1	2.5	0.9509	3.50	99.2
		NO		-0.1	2.5	1.0015	3.50	
N Lanarkshire Croy	18-Jul	NOx	ayktcju8	-0.7	2.4	0.7424	3.50	99.8
		NO		-0.1	2.4	0.7547	3.50	
N Lanarkshire Kenilworth Road	19-Jul	NOx	huk14100092	-1.1	2.5	0.9867	3.79	100.7
		NO		-0.3	2.5	1.0274	3.82	
N Lanarkshire Kirkshaws	16-Jul	NOx	huk15020066	-0.6	2.5	0.9988	3.50	98.8
		NO		-0.1	2.5	1.0253	3.50	
N Lanarkshire Motherwell	18-Jul	NOx	ypb4f-s4u	0.1	2.5	1.0017	3.50	98.6
		NO		0.3	2.5	1.0360	3.50	
N Lanarkshire New Edinburgh Road	22-Jul	NOx	hil10070072	-1.4	2.5	0.9760	3.50	100.5
		NO		-0.6	2.5	1.0037	3.50	
N Lanarkshire Shawhead Coatbridge	16-Jul	NOx	7nhskhbc	-0.7	2.5	0.9916	3.50	101.7
		NO		-0.4	2.5	1.0186	3.50	
N Lanarkshire Sunnyside Road	19-Jul	NOx	huk14070019	0.1	2.5	0.9772	3.50	101.3
		NO		0.1	2.5	1.0122	3.50	

PM10 analysers

Station	Date of audit	Analyser Serial no	Calculated ko	Uncertainty %	Total flow	Uncertainty %	Main flow	Uncertainty %
N Lanarkshire Chapelhall	18-Jul	8323			4.73	2.2		2.2
N Lanarkshire Coatbridge Whifflet	19-Jul	140ab217699710	12974	1.0	2.83	2.2		2.2
N Lanarkshire Croy	18-Jul	9552			4.77	2.2		2.2
N Lanarkshire Kenilworth Road	19-Jul	h11772			16.38	2.2		2.2
N Lanarkshire Kirkshaws	16-Jul	9553			4.80	2.2		2.2
N Lanarkshire Motherwell	18-Jul	9551			4.75	2.2		2.2
N Lanarkshire New Edinburgh Road	22-Jul	h18029			16.69	2.2		2.2
N Lanarkshire Shawhead Coatbridge	16-Jul	9550			4.68	2.2		2.2
N Lanarkshire Sunnyside Road	19-Jul	h11774			16.16	2.2		2.2

PM2.5 analysers

Station	Date of audit	Analyser Serial no	Calculated ko	Uncertainty %	Total flow	Uncertainty %	Main flow	Uncertainty %
N Lanarkshire Chapelhall	18-Jul	8323			4.73	2.2		2.2
N Lanarkshire Croy	18-Jul	9552			4.77	2.2		2.2
N Lanarkshire Kirkshaws	16-Jul	9553			4.80	2.2		2.2
N Lanarkshire Motherwell	18-Jul	9551			4.75	2.2		2.2
N Lanarkshire Shawhead Coatbridge	16-Jul	9550			4.68	2.2		2.2



# CERTIFICATE OF CALIBRATION



Page 3 of 3

Date of issue: 20 May 20  
Certificate Number: 4956  
Ricardo Energy & Environment ID: ED11194 / 4956

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NO<sub>x</sub> analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and  $k_0$  (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are given in ppb (parts per billion) mole fractions or ppm (parts per million) mole fractions.

<sup>1</sup> The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

<sup>2</sup> The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported concentration units (ppb for NO, NO<sub>x</sub>, SO<sub>2</sub>, O<sub>3</sub> and ppm for CO. Where 1ppm = 1000ppb). It should be used in conjunction with the zero response. A corrected concentration is calculated using the following equation:

$$\text{Concentration} = F(\text{Output} - \text{Zero Response})$$

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser

Zero Response = Zero Response provided on this certificate

<sup>3</sup> Converter eff. is the measured efficiency of the NO<sub>2</sub> to NO converter within the oxides of nitrogen analyser under test.

<sup>4</sup> The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are  $\text{l}\cdot\text{min}^{-1}$ , reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

<sup>5</sup> The calculated  $k_0$  value (specifically for TEOM analysers) is the calculated  $k_0$  spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified value of  $k_0$ .

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.