



CERTIFICATE OF CALIBRATION

Ricardo Energy and Environment, Gemini Building, Fermi Avenue Harwell,
Didcot, Oxfordshire OX11 0QR. Telephone 01235 753692



Approved Signatories:

- | | | | |
|--------------------------|-------------|-------------------------------------|------------|
| <input type="checkbox"/> | S. Eaton | <input type="checkbox"/> | B Stacey |
| <input type="checkbox"/> | D Hector | <input type="checkbox"/> | S Stratton |
| <input type="checkbox"/> | N Rand | <input type="checkbox"/> | S Telfer |
| <input type="checkbox"/> | E Marshall- | <input checked="" type="checkbox"/> | S Gray |
| <input type="checkbox"/> | Padkin | | |
| <input type="checkbox"/> | B Davies | | |

Signed:

Date of issue:

25 Apr 18

Certificate Number:

3937

Customer Name and Address:

Scottish Government
Water, Air, Soils and Flooding Division
Environmental Quality Directorate
Scottish Government
Victoria Quay
Edinburgh
EH6 6QQ

Description:

Calibration factors for the air monitoring stations at
South Lanarkshire Council

Ricardo Energy & Environment ID:

ED61598/3937

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k=2$ providing a level of confidence of approximately 95% The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory

Ricardo Energy & Environment

Head Office
Gemini Building,
Fermi Avenue,
Harwell,
Oxon
OX11 0QR

Tel: +44 (0)1235 753 000

Registered office

Shoreham Technical Centre
Shoreham-by-Sea
West Sussex
BN43 5FG

Registered in England No.
08229264

VAT Registration No.
GB 212 8365 24



CERTIFICATE OF CALIBRATION



Date of issue: 25 Apr 18
 Certificate Number: 3937
 Ricardo Energy & Environment ID: ED61598/3937

South Lanarkshire Council

NOx analysers

Station	Date of Audit	Species	Analyser Serial no	Zero Response ¹	Zero uncertainty ppb	Calibration Factor ²	Factor uncertainty %	Converter eff. (%) ³
South Lanarkshire East Kilbride	17-Jul-17	NOx	CM07460075	4.3	2.6	1.1641	3.50	99.2
		NO		4.2	2.6	1.1612	3.50	
South Lanarkshire Hamilton	18-Jul-17	NOx	CM07460073	-22.8	2.5	0.9557	3.50	98.2
		NO		-22.8	2.5	0.9536	3.50	
South Lanarkshire Lanark	18-Jul-17	NOx	CM10020067	0.3	2.5	1.0235	3.50	101.3
		NO		-0.1	2.5	1.0226	3.50	
South Lanarkshire Rutherglen	17-Jul-17	NOx	CM07460076	-5.6	2.6	1.0650	3.50	99.3
		NO		-7.0	2.5	1.0616	3.50	
South Lanarkshire Uddingston	17-Jul-17	NOx	CM10020068	-0.8	2.6	1.1316	3.50	100.0
		NO		-1.3	2.6	1.1333	3.50	
South Lanarkshire Cambuslang	18-Jul-17	NOx	1152590008	-1.6	3.1	2.0171	3.50	99.0
		NO		-1.3	3.1	2.0244	3.50	
South Lanarkshire Raith Interchange	17-Jul-17	NOx	CM10220001	-1.0	2.6	1.1197	3.50	98.7
		NO		-0.4	2.6	1.1161	3.50	

PM10 analysers

Station	Date of audit	Analyser Serial no	Calculated ko	Uncertainty %	Total flow	Uncertainty %	Main flow	Uncertainty %
South Lanarkshire Cambuslang	18-Jul-17	8256			4.83	2.2		2.2
South Lanarkshire East Kilbride	17-Jul-17	8257			4.60	2.2		2.2
South Lanarkshire Lanark	18-Jul-17	6248			4.82	2.2		2.2
South Lanarkshire Rutherglen	17-Jul-17	8140			4.77	2.2		2.2
South Lanarkshire Uddingston	17-Jul-17	6247			4.70	2.2		2.2
South Lanarkshire Raith Interchange	17-Jul-17	27186	15263	1.0	15.31	2.2	3.06	2.2



CERTIFICATE OF CALIBRATION



Date of issue: 25 Apr 18
Certificate Number: 3937
Ricardo Energy & Environment ID: ED61598/3937

The gaseous ambient analysers listed above have been tested for zero response, calibration factor, linearity and converter efficiency (NO_x analysers) by documented methods. The factors have been calculated using certified gas standards. The particulate analysers listed above have been tested for sample flow rates and k_0 (where appropriate) by documented methods. Note that the test results are valid on the day of test only, as analyser drift over time cannot be quantified. All results for gaseous species are given in ppb (parts per billion) mole fractions or ppm (parts per million) mole fractions.

¹ The zero response is the zero reading on the data logging system of the analyser when audit zero gas was introduced to the analysers under test.

² The calibration factor is the multiplying factor required to scale the reading on the data logging system of the analyser into reported concentration units (ppb for NO, NO_x, SO₂, O₃ and ppm for CO. Where 1ppm = 1000ppb). It should be used in conjunction with the zero response. A corrected concentration is calculated using the following equation:

$$\text{Concentration} = F(\text{Output} - \text{Zero Response})$$

Where F = Calibration Factor provided on this certificate

Output = Reading on the data logging system of the analyser

Zero Response = Zero Response provided on this certificate

³ Converter eff. is the measured efficiency of the NO₂ to NO converter within the oxides of nitrogen analyser under test.

⁴ The measured main flow rate (where this is applicable) is the flow rate through the sensor unit of the TEOM particulate analyser under test. The measured total flow rate is the total flow rate through the particulate analyser under test. Units of flow are l.min⁻¹, reported at prevailing ambient conditions unless otherwise specified. Where flow rates are highlighted in bold, it indicates that measurements were not made at the analyser sample inlet. These measurements therefore may not accurately reflect analyser performance in normal operation.

⁵ The calculated k_0 value (specifically for TEOM analysers) is the calculated k_0 spring constant based on tests undertaken with filters of known weight. The % deviation indicates the closeness of the calculated result to the manufacturer's specified value of k_0 .

The calibration results shaded are those that fall within our scope of accreditation, all other results on this certificate are not UKAS accredited, but have been included for completeness.